



### Instructions For Use

## LightMix<sup>®</sup> Modular *Listeria monocytogenes*

530

Cat.-No. 53-0642-96

Roche SAP n° 07 413 700 001

Kit with reagents for 96 PCR reactions 20 µl for detection of Giardia [lyophilized]

### 1. Content, Storage and Expiry

#### Storage at Arrival:

1 Vial yellow cap 96 reactions *Listeria* (lyophilized)

Store cooled or at ambient temperature

1 Vial black cap Positive Control (32,000 copies, lyophilized)

Do **not** freeze the lyophilized reagents.

- Lyophilized kits are stable for at least 6 months (4°C to 25°C in the dark). See lot-specific expiry date.
- Dissolved reagents are stable for at least 2 weeks if stored protected from light and cooled (4°C).
- Dissolved reagents can be stored long-term at -20°C (within expiry). Avoid multiple freeze-thaw cycles.
- Dissolved positive controls must be stored at -20°C. Avoid multiple freeze-thaw cycles.

### 2. Additional Reagents required

LightCycler<sup>®</sup> Multiplex DNA Master

Cat.-No. 07 339 585 001

or Roche LightCycler<sup>®</sup> 480 Probes Master (no instructions included)

Cat.-No. 04 707 494 001

### 3. Introduction

Meningitis is an inflammation of the membranes in the brain caused by infection with viruses, bacteria or other microorganisms. For newborn some cases are caused by *Listeria monocytogenes*, a Gram-positive bacterium better known for foodborne infections.

### 4. Description

A 89 bp fragment of the *hlyA* gene of *L. monocytogenes* is amplified with specific primers and detected with a FAM labeled hydrolysis probe.

### 5. Specification

This assay detects 10 genome equivalent copies or less per reaction (plasmid DNA dilution).

### 6. Sample Material and Extraction

Typical samples are cerebrospinal fluid (CSF). See ModularDx Document **Extraction Protocols**.

### 7. Material Safety Data (MSDS)

According to OSHA 29CFR1910.1200, Australia [NOHSC:1005, 1008 (1999)] and the EU Directives 67/548/EC and 1999/45/EC any products which do not contain more than 1% of a component classified as hazardous or classified as carcinogenic do not require a Material Safety Data Sheet (MSDS).

Product is not hazardous, not toxic, not IATA-restricted. Product is not from human, animal or plant origin. Product contains synthetic oligonucleotide primers and probes.



## 8. Instructions for Use

- Instrument programming see document **ModularDx Programming**
- Color Compensation see instructions in **40-0320 Universal Color Compensation Hexaplex**
- Pipetting instructions multiplex PCR see **ModularDx Multiplex**

### 8.1. Programming Roche 480 Instruments

See the Instrument operator's manual for details. Start programming before preparing the solutions. The protocol consists of three program steps:

- 1: Denaturation: sample denaturation and enzyme activation
- 2: Cycling: PCR-amplification
- 3: Cooling: cooling the instrument

<b>Detection Format 530 Channel</b>	<b>Set Quant Factor 10</b> (default setting is 1)
LightCycler® 480 Instrument:	483-533
LightCycler® 480 II Instrument:	465-510
cobas z 480 Analyzer (open channel):	465-510

Program Step:	RT Step*	Denaturation	Cycling			Cooling
<b>Parameter</b>						
Analysis Mode	None	None	Quantification mode			None
Cycles	1	1	45			1
Target [°C]	55	95	95	60	72	40
Hold [hh:mm:ss]	00:05:00	00:05:00	00:00:05	00:00:15	00:00:15	00:00:30
Ramp Rate [°C/s] <b>96</b>	4.4	4.4	4.4	2.2	4.4	1.5
Ramp Rate [°C/s] <b>384</b>	4.6	4.6	4.6	2.4	4.6	2.0
Acquisition Mode	None	None	None	Single	None	None

\* optional to combine with 1-Step RT-PCR

Table 1

### 8.2. Experimental Protocol

- **Sample material:** Use aqueous nucleic acid preparations (e.g. 'High Pure PCR Template Preparation Kit').
- **Negative control:** Always run at least one no-template control (NTC) - replace the template NA with water.
- **Positive control:** Run a positive control - replace the template NA with the provided Positive Control.

For an increased sensitivity use 10 µl sample per 20 µl reaction, in case that inhibition is likely to occur, e.g. extracts obtained from fecal samples, use 5 µl. For 10 µl reactions in 384 well plates use 5 µl /2.5 µl.

#### 8.2.1. Preparation of Parameter-Specific Reagents (PSR, 96 reactions):

One reagent vial with a **yellow** cap contains all primers and probe to run 96+ LightCycler® reactions.

**Add 50 µl** PCR-grade water to each reagent vial, mix the solution (vortex) and spin down. For robotic pipetting the volume can be extended to 55 µl (signals will decrease by 10-20%).

► **Use 0.5 µl** reagent for a 20 µl PCR reaction.

#### 8.2.2. Preparation of the Positive Control

**Add 160 µl** RNase/DNase-free Tris buffer or water to the vial with the **black** cap, for 10 µl sample add **320 µl**. Mix the solution by pipetting up and down 10 times. If vortexing spin down to collect the solution.

**Notes:** Opening of this vial may cause contaminations of the work-space (aerosol). Use of Tris buffer pH 8.0-8.5 increases the long-term stability in solution. Store dissolved controls frozen.

► **Use 5 µl** positive control (≈ 1,000 copies) for a 20 µl PCR reaction (or 10 µl if using 10 µl sample).

**Note:** The Positive Control can yield late positive results for *E.coli* assays such as uidA, tnaA, or agaB.

### 8.2.3. Preparation of the Reaction Mix

Multiply volumes by the number of reactions plus controls and one reserve and prepare in a cooled tube:

For use with the Roche LightCycler® Multiplex DNA Master		
for 5 µl extract	Component	10 µl extract
10.5 µl	<b>Water</b> , PCR-grade (colorless cap, provided with the Roche Master kit)	5.5 µl
0.5 µl	<b>Reagent</b> mix (parameter specific reagents containing primers and probes)	0.5 µl
--	Control Reaction and additional assays (Multiplex PCR)	--
4.0 µl	<b>Roche Master</b> (see Roche manual)	4.0 µl
<b>15.0 µl</b>	<b>Volume of Reaction Mix</b>	<b>10.0 µl</b>


Table 2

Mix gently, spin down and **transfer 15 µl (10 µl)** per well.

**Add 5 µl (10 µl)** of sample or control to each well for a final reaction volume of 20 µl. Seal plate and centrifuge.

**Start run**

### 9. Typical Results (Data from LightCycler® 480 II system)



Dilution row 1E6 to 10 copies / reaction

Figure 1

### 10. Reading the Results

Perform data analysis as described in the operator's manual. For multiplex assays select the color compensation. We recommend using the Second Derivative Maximum method (Automated (F'' max)). View results in the FAM channel. The negative control (NTC) must show no signal.

Channel 530 (sample)	Channel 660 Control Reaction	Channel 530 NTC Control	Result
No amplification	Detectable	Negative	Not detectable
Amplification Cp < 37 <sup>+</sup>	Not relevant	Negative	Listeria Positive
No amplification	Not detectable	Not relevant	PCR failure Repeat
Amplification signal	Not relevant	Positive	Contamination Repeat

**Notes:** cobas z 480 Analyzer signal levels are about 50% compared to LightCycler® 480 II results.  
<sup>+</sup> Recommendation : Define the cut-off 2-4 cycles higher than observed for 10 copies.

### 11. References

Application of 5'-nuclease PCR for quantitative detection of Listeria monocytogenes in pure cultures, water, skim milk, and unpasteurized whole milk. Nogva et al. Appl Environ Microbiol 66 (2000) 4266-71

## 12. Multiplex PCR Compatibility (Meningitis Panel)

This Listeria assay can be combined with other assays up to 6plex reactions including an internal control (IC) or a spiked extraction control (for example PhHV) as depicted below :

### Meningitis Multiplex PCR and Instrument Compatibility

Color Compensation 40-0320 is mandatory for Multiplex PCR


500	530	580	610	640	660
	L.mono				PhHV
	L.mono	S.pneumo	S.aga		PhHV
	L.mono	S.pneumo	N.men		PhHV
	L.mono	S.pneumo	S.aga	H.influenz	PhHV
	L.mono	S.pneumo	N.men	H.influenz	PhHV
E.coli	L.mono	S.pneumo	S.aga	H.influenz	PhHV
E.coli	L.mono	S.pneumo	N.men	H.influenz	PhHV

480 II	z 480	LC96	LC2.0	Nano
X	X	X	X	X
X	X	X		
X	X	X		
X	X			
X	X			
X				
X				

Table 3

## 13. Version History

V140404	Release version	2014-04-14
V140909	Editorial changes	2014-09-09
V151001	2015 protocol Multiplex Master, 5/10 µl extract and 60°C acquisition	2015-10-01
V160101	Positive control may be visible in ModularDx kits for <i>E.coli</i> targets	2016-01-07
V160313	1. Storage of controls, 8.2.2 buffer, 8.2.3 wording	2016-06-10

Certificate of Analysis (CoA)							
Lot n° Expiry :							
<b>Dilution</b>	<b>1E6</b>	<b>1E5</b>	<b>1E4</b>	<b>PC</b>	<b>1E2</b>	<b>1E1</b>	<b>passed</b>
<b>Cp range</b>	19-21	22-24	25-27	29-31	32-34	35-37	
<b>Measured Signal level</b>	40-60						
<b>Measured</b>							
<b>Negatives</b>	<b>10/10</b>						✓
<b>Note:</b> Cp (crossing point) values collected with pDNA (single target PCR). Fluorescence (FL) levels depend on instrument settings and may vary. The Cp values will vary from instrument to instrument by up to 2 cycles, while the distance between two dilution steps should be relative constant ( $\Delta Cp$ ).							
QC Acceptance Date:				YYYYMMDD			
We, the undersigned, certify that the product designated above has been obtained in accordance with the rules of production and quality control.							
Name(s) :							

**TIB MOLBIOL** Syntheselabor GmbH | Eresburgstr. 22-23 | D-12103 Berlin | Germany  
 Tel. +49 30 78 79 94 55 | FAX +49 78 79 94 99 | dna@tib-molbiol.de | WWW.TIB-MOLBIOL.COM  
 Geschäftsführer (CEO): Olfert Landt | Register HRB 93163 B | Registergericht Berlin Charlottenburg

Distributed by Roche - [www.lifescience.roche.com](http://www.lifescience.roche.com)